



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Clinical Reference Laboratory-Gen Lab*, using *Roche DBS on Cobas c513* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: *University of Missouri SRL#9*

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 62634601	Application: HbA1c% = (HbA1c/Hb) x 91.5 + 2.15
Reagent Lot: 67251001, 70526601	Calibrator Assigned Values: 0.240, 0.552, 1.050, 1.435, 1.969, 2.782 g/dL HbA1c; 2.33, 12.53 g/dL Hb	

Date of Certification: August 1, 2023

Certification Expires: August 1, 2024

A handwritten signature in black ink, appearing to read "David Sacks".

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read "Randie R Little PhD".

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read "Sharon Anulf".

SRL director/ supervisor