



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Clinical Reference Laboratory – Gen Lab*, using *Roche Cobas c513* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 559759	Application: HbA1c% = (HbA1c/Hb) x 91.5 + 2.15
Reagent Lot: 645100, 671531	Calibrator Assigned Values: 0.246, 0.560, 1.063, 1.454, 1.995, 2.819 g/dL HbA1c; 2.33, 12.53 g/dL Hb	

Date of Certification: June 1, 2023

Certification Expires: June 1, 2024

Daniel Sacks

NGSP Steering Committee Chair

Randie R Little PhD

NGSP Network Coordinator

Sharon Amalf

SRL director/ supervisor