



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory – Olathe**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 559759	Application: HbA1c%=(HbA1c/Hb)x91.5+2.15
Reagent Lot: 629580, 646940, 698206	Calibrator Assigned Values: 0.254, 0.573, 1.079, 1.473, 2.017, 2.850 g/dL HbA1c; 2.33, 12.53 g/dL Hb	

Date of Certification: September 1, 2023

Certification Expires: September 1, 2024

Daniel Sacks
NGSP Steering Committee Chair

Randie R Little PhD.
NGSP Network Coordinator

Sharon Amalfi
SRL director/ supervisor